

CERTIFICATE OF CONFORMANCE

Droplet Measurement Technologies, Inc.

certifies that the

**Cloud Condensation Nuclei Counter-Single Column
(CCN-100/AAA-0024)
S/N: 0604-001**


has received factory optical cleaning, optical alignment, and calibrations. The instrument is certified to meet all factory calibration tolerances in accordance with DMT standards as of the date of this certificate.

Certificate Date.....4/27/2022

Sales Order.....1706391

Expiration Month11/2022

Calibration by:



*D Johnson
CCN Product Engineer*

Verified by:



*J Klotsche
Vice President Finance/Accounting*

Atmospheric Particle Spectrometers are used in very harsh environments and require routine cleaning and calibration verification during campaigns. The user is responsible for routine maintenance. DMT advises a full factory optical cleaning, optical alignment and calibration adjustment once per year.



2400 Trade Center Avenue
Longmont, CO USA 80305
Tel: 303-440-5576

www.dropletmeasurement.com
customer-contact@dropletmeasurement.com